

Specification Sheet

IP2036, Rev. BA

July 2013

Level

Mobrey MSP-RTP

External Temperature Sensor

Part No : MSP-RTP

For use with ultrasonic transmitters MSP400RH and MSP900GH



Introduction

For transmitter installations where the integral temperature sensor is not representative of the air / medium temperature in the process, the use of the external temperature sensor improves accuracy.

Sensor Specification

316 Stainless steel body with two locknuts, M8 thread, 6 mm maximum panel thickness. Sensor working temperature -40 to 85 °C, cable working temperature -20 to 80 °C. Cable 5 m long, single twisted pair.

Note :

When connected to a MSP900GH intrinsically safe transmitter, this sensor is classified as 'simple apparatus' according to EN 50020 paragraph 5.4. Temperature class T6, $-20\text{ °C} \leq T_a \leq 80\text{ °C}$

Level

Installation

1. Connect the two sensor wires to terminals 7 and 8 of the transmitter; polarity is not important. See diagram below; also refer to Transmitter Installation Manual.

Ensure use of a suitably IP rated cable gland to maintain the IP rating of the transmitter.

2. Mount the sensor between the transmitter and the liquid level, to measure the air / medium temperature at a representative point.


3. Avoid allowing the sensor to become submerged.

4. Do not mount the sensor near a heat source or in direct sunlight.

5. For the optimum response to temperature change, do not mount directly onto a large thermal mass.

6. To avoid false echoes, do not mount the sensor directly in the path of the transmitter beam.

7. The transmitter will automatically detect and enable the sensor. Fine adjustments to the temperature calibration may be made; refer to the Installation Manual supplied with the transmitter.



Wiring Diagram
External temperature sensor

*The Emerson logo is a trademark and service mark of Emerson Electric Co.
Rosemount and the Rosemount logotype are registered trademarks of Rosemount Inc.
Mobrey is a registered trademark of Mobrey Ltd.
All other marks are the property of their respective owners.*

© 2013 Mobrey Ltd. All rights reserved.

**Emerson Process Management
Mobrey Ltd.**
158 Edinburgh Avenue
Slough, Berks, SL1 4UE, UK
T +44 (0)1753 756600
F +44 (0)1753 823589
www.mobrey.com

**Emerson Process Management
Rosemount Measurement**
8200 Market Boulevard
Chanhassen, MN 55317 USA
T (USA) 1 800 999 9307
T (Intl) +1 952 906 8888
F +1 952 906 8889